

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 055581	(X2) MULTIPLE CONSTRUCTION A. Building B. Wing	(X3) DATE SURVEY COMPLETED 12/04/2024
NAME OF PROVIDER OR SUPPLIER Jurupa Hills Post Acute		STREET ADDRESS, CITY, STATE, ZIP CODE 6401 33rd Street. Riverside, CA 92509	

For information on the nursing home's plan to correct this deficiency, please contact the nursing home or the state survey agency.

(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (Each deficiency must be preceded by full regulatory or LSC identifying information)
<p>F 0908</p> <p>Level of Harm - Minimal harm or potential for actual harm</p> <p>Residents Affected - Few</p>	<p>Keep all essential equipment working safely.</p> <p>**NOTE- TERMS IN BRACKETS HAVE BEEN EDITED TO PROTECT CONFIDENTIALITY** 46509</p> <p>Based on observation, interview, and record review, the facility failed to ensure one (Resident A) of four sampled residents' wheelchair was maintained in a safe and operable condition.</p> <p>This failure had the potential to cause injury to Resident A when he is using a wheelchair with two broken wheel brakes.</p> <p>Findings:</p> <p>On November 15, 2024, at 5:30 p.m., an unannounced visit to the facility was conducted to investigate three complaints.</p> <p>On November 18, 2024, at 12:30 p.m., a review of Resident A's medical record was conducted. Resident A was admitted to the facility on [DATE], with diagnoses which included Hemiplegia and Hemiparesis (muscle weakness or partial paralysis on one side of the body).</p> <p>Resident A's care plan, dated July 9, 2024, indicated Resident A had .decreased functional mobility .</p> <p>On November 18, 2024, at 1:37 p.m., an observation and concurrent interview was conducted with Resident A. Resident A stated he was admitted to the facility after he had a stroke. Resident A stated the locks on his wheelchair are not working, and the wheelchair is a little small for him. Observed Resident A attempt to apply the brakes on his wheelchair, he was unable to lock the wheels in place, both brakes were broken. Observed Resident A in his wheelchair, Resident A's hips were pushed against the sides of the wheelchair seat, unable to move around and reposition himself.</p> <p>On November 18, 2024, at 1:50 p.m., and interview was conducted with the Director of Rehabilitation Services (DRS). The DRS stated all residents should be using a wheelchair in working order, and the brakes should work. The DRS stated Resident A will need a wheelchair with brakes that are operatable, this is a safety concern, especially if Resident A is transferring from bed to the wheelchair, or the other way around. The DRS noted Resident A's wheelchair seat was small for Resident A's size, we will get Resident A a new chair that is larger and has working brakes, his current wheelchair is safety concern.</p> <p>(continued on next page)</p>

Any deficiency statement ending with an asterisk (*) denotes a deficiency which the institution may be excused from correcting providing it is determined that other safeguards provide sufficient protection to the patients. (See instructions.) Except for nursing homes, the findings stated above are disclosable 90 days following the date of survey whether or not a plan of correction is provided. For nursing homes, the above findings and plans of correction are disclosable 14 days following the date these documents are made available to the facility. If deficiencies are cited, an approved plan of correction is requisite to continued program participation.

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE	TITLE	(X6) DATE
---	-------	-----------

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 055581	(X2) MULTIPLE CONSTRUCTION A. Building B. Wing	(X3) DATE SURVEY COMPLETED 12/04/2024
NAME OF PROVIDER OR SUPPLIER Jurupa Hills Post Acute		STREET ADDRESS, CITY, STATE, ZIP CODE 6401 33rd Street. Riverside, CA 92509	
For information on the nursing home's plan to correct this deficiency, please contact the nursing home or the state survey agency.			
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (Each deficiency must be preceded by full regulatory or LSC identifying information)		
<p>F 0908</p> <p>Level of Harm - Minimal harm or potential for actual harm</p> <p>Residents Affected - Few</p>	<p>A review of the facility's policy titled Maintenance Service , dated December 2009, indicated .Maintenance service shall be provided to .equipment .maintenance department is responsible for maintaining the . equipment in a safe and operable manner at all times .Maintenance personnel shall follow established safety regulations to ensure safety and well-being of all .</p>		