

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER:  155549	(X2) MULTIPLE CONSTRUCTION A. Building B. Wing	(X3) DATE SURVEY COMPLETED  04/16/2025
NAME OF PROVIDER OR SUPPLIER  Envive of Muncie		STREET ADDRESS, CITY, STATE, ZIP CODE  7524 E Jackson Street Muncie, IN 47302	
For information on the nursing home's plan to correct this deficiency, please contact the nursing home or the state survey agency.			
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (Each deficiency must be preceded by full regulatory or LSC identifying information)		
<p>F 0880</p> <p>Level of Harm - Minimal harm or potential for actual harm</p> <p>Residents Affected - Some</p>	<p>Provide and implement an infection prevention and control program.</p> <p>32663</p> <p>Based on observation, interview, and record review, the facility failed to ensure staff (LPN 1) followed the facility cleaning protocol for resident glucometers to reduce the risk of contamination and spread of infection for 5 of 5 residents reviewed for infection control during medication administration. (Resident B, H, J, K, and L)</p> <p>Findings include:</p> <p>During a medication administration observation on 4/15/25 at 5:02 a.m., the following infection control concerns were observed:</p> <p>LPN 1 removed the Resident B's glucometer from a plastic bag in the medication cart. The glucometer was placed directly on the medication cart while LPN 1 gathered supplies and medications. Upon entering the resident's room, LPN 1 placed the glucometer on the over the bed table. After performing the blood sugar check, LPN 1 returned to the medication cart and placed the glucometer directly on the cart. She then returned it to the plastic bag and secured it to the medication cart. The glucometer was not disinfected during the observation. The top of the medication cart was not cleaned/disinfected during the observation. The resident's over bed table was not cleaned prior to placing the glucometer on the table.</p> <p>LPN 1 removed the Resident H's glucometer from a plastic bag in the medication cart. The glucometer was placed directly on the medication cart while LPN 1 gathered supplies and medications. Upon entering the resident's room, LPN 1 placed the glucometer on the over the bed table. After performing the blood sugar check, LPN 1 returned to the medication cart and placed the glucometer directly on the cart. She then returned it to the plastic bag and secured it to the medication cart. The glucometer was not disinfected during the observation. The top of the medication cart was not cleaned/disinfected during the observation. The resident's over bed table was not cleaned prior to placing the glucometer on the table.</p> <p>(continued on next page)</p>		

Any deficiency statement ending with an asterisk (\*) denotes a deficiency which the institution may be excused from correcting providing it is determined that other safeguards provide sufficient protection to the patients. (See instructions.) Except for nursing homes, the findings stated above are disclosable 90 days following the date of survey whether or not a plan of correction is provided. For nursing homes, the above findings and plans of correction are disclosable 14 days following the date these documents are made available to the facility. If deficiencies are cited, an approved plan of correction is requisite to continued program participation.

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE	TITLE	(X6) DATE
---	-------	-----------

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER:  155549	(X2) MULTIPLE CONSTRUCTION A. Building B. Wing	(X3) DATE SURVEY COMPLETED  04/16/2025
NAME OF PROVIDER OR SUPPLIER  Envive of Muncie		STREET ADDRESS, CITY, STATE, ZIP CODE  7524 E Jackson Street Muncie, IN 47302	
For information on the nursing home's plan to correct this deficiency, please contact the nursing home or the state survey agency.			
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (Each deficiency must be preceded by full regulatory or LSC identifying information)		
<p>F 0880</p> <p>Level of Harm - Minimal harm or potential for actual harm</p> <p>Residents Affected - Some</p>	<p>LPN 1 removed the Resident J's glucometer from a plastic bag in the medication cart. The glucometer was placed directly on the medication cart while LPN 1 gathered supplies and medications. Upon entering the resident's room, LPN 1 placed the glucometer on the over the bed table. After performing the blood sugar check, LPN 1 returned to the medication cart and placed the glucometer directly on the cart. She then returned it to the plastic bag and secured it to the medication cart. The glucometer was not disinfected during the observation. The top of the medication cart was not cleaned/disinfected during the observation. The resident's over bed table was not cleaned prior to placing the glucometer on the table.</p> <p>LPN 1 removed the Resident K's glucometer from a plastic bag in the medication cart. The glucometer was placed directly on the medication cart while LPN 1 gathered supplies and medications. Upon entering the resident's room, LPN 1 placed the glucometer on the over the bed table. After performing the blood sugar check, LPN 1 returned to the medication cart and placed the glucometer directly on the cart. She then returned it to the plastic bag and secured it to the medication cart. The glucometer was not disinfected during the observation. The top of the medication cart was not cleaned/disinfected during the observation. The resident's over bed table was not cleaned prior to placing the glucometer on the table.</p> <p>LPN 1 removed the Resident L's glucometer from a plastic bag in the medication cart. The glucometer was placed directly on the medication cart while LPN 1 gathered supplies and medications. Upon entering the resident's room, LPN 1 placed the glucometer on the over the bed table. After performing the blood sugar check, LPN 1 returned to the medication cart and placed the glucometer directly on the cart. She then returned it to the plastic bag and secured it to the medication cart. The glucometer was not disinfected during the observation. The top of the medication cart was not cleaned/disinfected during the observation. The resident's over bed table was not cleaned prior to placing the glucometer on the table.</p> <p>During the medication administration observation, LPN 1 indicated each resident had their own glucometer.</p> <p>During an interview on 4/15/25 at 9:03 a.m., RN 2 indicated glucometers were to be cleaned and allowed to dry before placing them back into the plastic storage bag.</p> <p>During an interview on 4/16/25 at 1:27 p.m. , the DON indicated resident glucometers were to be cleaned/disinfected prior to returning them to the medication cart.</p> <p>Review of the manufacture's cleaning guidelines, provided by the Regional Nurse on 4/16/25 at 1:13 p.m., indicated the following: .The disinfecting procedure is needed to prevent the transmission of bloodborne pathogens .</p> <p>Review of a current policy, dated 8/2024 and titled Obtaining a Fingertick Glucose Level, was provided by the Administrator on 4/15/25 at 10:17 a.m. and indicated the following: Steps in the Procedure 18. Clean and disinfect reusable equipment between uses according to the manufacture's instructions and current infection control standards of practice.</p> <p>This citation relates to Complaint IN00456768.</p> <p>3.1-18(l)</p>		