

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 245186	(X2) MULTIPLE CONSTRUCTION A. Building B. Wing	(X3) DATE SURVEY COMPLETED 04/04/2024
NAME OF PROVIDER OR SUPPLIER The Villas at Brookview		STREET ADDRESS, CITY, STATE, ZIP CODE 7505 Country Club Drive Golden Valley, MN 55427	

For information on the nursing home's plan to correct this deficiency, please contact the nursing home or the state survey agency.

(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (Each deficiency must be preceded by full regulatory or LSC identifying information)
<p>F 0880</p> <p>Level of Harm - Minimal harm or potential for actual harm</p> <p>Residents Affected - Few</p>	<p>Provide and implement an infection prevention and control program.</p> <p>27955</p> <p>Based observation, interview and document review, the facility failed to ensure community use glucometer were properly cleaned and disinfected between resident use for 1 of 2 glucometers on the second floor.</p> <p>Findings include:</p> <p>During an observation on 4/4/24 at 10:36 a.m., licensed practical nurse (LPN)-A gathered the supplies to perform a blood sugar reading for R12. LPN-A went to R12's room. LPN-A inserted the test strip, donned gloves, wiped R12's finger with alcohol wipe and used the lancet to poke the finger. LPN-A applied the drop of blood to the test strip and got a reading of 193. LPN-A removed the gloves and took the community glucometer (more than one resident utilizes the glucometer for testing. If blood glucose meters must be shared, the device should be cleaned and disinfected after every use, per manufacturer's instructions, to prevent carry-over of blood and infectious agents) and sharp out of the room. LPN-A placed the sharp in the sharps container. LPN-A took an alcohol wipe and wiped the community glucometer. LPN-A put the community glucometer in the medication cart and locked it.</p> <p>On 4/4/24 at 10:42 a.m., LPN-A stated she did use an alcohol wipe to clean the community glucometer after each use.</p> <p>On 4/4/24 at 10:51 a.m., registered nurse (RN)-A stated the glucometers were community use. RN-A stated that an alcohol wipe was not to be used to clean the community glucometers. RN-A stated a purple top Sani wipe was to be used to clean the community glucometer.</p> <p>On 4/12/24 at 11:02 a.m. the director of nursing (DON) stated we clean community glucometers with the purple Sani wipe. The DON stated there is a two-minute kill time (the surface should remain visibly wet for two minutes after the solution has been applied to eliminate or kill specific bacteria, virus, or fungi) then could be put away for the next resident use. The DON stated and alcohol wipe should not be used to clean the community glucometer.</p> <p>Assure Platinum blood glucose monitoring system instruction manual indicated cleaning and disinfecting guidelines included:</p> <p>-Cleaning and disinfecting could be completed by using a commercially available EPA (Environmental Protection Agency) registered disinfectant detergent or germicide wipe.</p> <p>(continued on next page)</p>

Any deficiency statement ending with an asterisk (*) denotes a deficiency which the institution may be excused from correcting providing it is determined that other safeguards provide sufficient protection to the patients. (See instructions.) Except for nursing homes, the findings stated above are disclosable 90 days following the date of survey whether or not a plan of correction is provided. For nursing homes, the above findings and plans of correction are disclosable 14 days following the date these documents are made available to the facility. If deficiencies are cited, an approved plan of correction is requisite to continued program participation.

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE	TITLE	(X6) DATE
---	-------	-----------

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION	(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 245186	(X2) MULTIPLE CONSTRUCTION A. Building B. Wing	(X3) DATE SURVEY COMPLETED 04/04/2024
NAME OF PROVIDER OR SUPPLIER The Villas at Brookview		STREET ADDRESS, CITY, STATE, ZIP CODE 7505 Country Club Drive Golden Valley, MN 55427	
For information on the nursing home's plan to correct this deficiency, please contact the nursing home or the state survey agency.			
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (Each deficiency must be preceded by full regulatory or LSC identifying information)		
<p>F 0880</p> <p>Level of Harm - Minimal harm or potential for actual harm</p> <p>Residents Affected - Few</p>	<p>-To use a wipe, remove from container and follow product label instructions to disinfect the meter. Take extreme care not to get liquid in the test strip and key code ports of the meter.</p> <p>-Many wipes act as both a cleaner and disinfectant, though if blood is visible present on the meter, two wipes must be used; use one wipe to clean and the second wipe to disinfect.</p>		